


ATLAS ITK STRIP SENSOR QUALITY ASSURANCE

Zhan Li

On behalf of ATLAS ITk group in China

CLHCP 2025, Henan

2025/11/2



Content


- ATLAS ITk Strip Sensor Quality Assurance Strategy
- CSNS Irradiation Site Qualification
- IHEP QA Site Qualification

ATLAS ITk Strip Sensor Quality Assurance

- To upgrade ATLAS inner detector, nearly 22,000 large area sensors will be produced over about five years by Hamamatsu Photonics K.K. (HPK).
- To monitor the production of ATLAS ITk sensor, the strategy of sensor quality assurance (QA) has been carried out.
- ATLAS ITk collaboration has prepared a detailed irradiation and test plan.

ITk sensor QA strategy

- Wafer:
 - Large-area MAIN sensor
 - The periphery remaining (the “halfmoons”) — test devices used for QA:
 - $1 \times 1 \text{ cm}^2$ miniature strip detectors (“Mini”)
 - large area ($8 \times 8 \text{ mm}^2$) diodes (“MD8”)
 - ATLAS test chip
- Mini used for CCE test, under $1.6 \times 10^{15} \text{ neq/cm}^2$ dose
- MD8 used for IV and CV test, focused on V_{FD}




Dicing scheme for the wafers (left), and detail of the QA-pieces design and size, “Testchip & MD8” (center) and “Mini & MD8” (right).

Planned irradiations

- All QA-pieces will be received at CERN first, then distributed to different irradiation sites.
- ITK strip sensors need to tolerate around an integrated total ionizing dose (TID) of 66 Mrad and a neutron-equivalent fluence of 1.6×10^{15} neq/cm²
- After irradiation, the QA pieces are sent to test sites for measurement.
- Could CSNS become an irradiation site?

Mini&MD8



Production flow for Mini&MD8 QA-pieces including irradiation and tests

China Spallation Neutron Source (CSNS)


- Located in Dongguan.
- First pulsed neutron source facility in China.
- It now includes:
 - a powerful linear accelerator and a rapid circling synchrotron
 - a target station and three Phase I neutron instruments





Aerial View of CSNS

Associated Proton Experiment Platform (APEP)

- Located at the end of the CSNS linear accelerator
- Proton beam energy 80 MeV
- 20 mm × 20 mm (flat top of flux) beam spot
- The simulated flux intensity at the sample location for a beam spot of this size is $3.04 \times 10^9 \text{ p/cm}^2/\text{s}$
- Take around 102.3 hours to reach $1.6 \times 10^{15} \text{ neq/cm}^2$





2025-11-1





CSNS irradiation site qualification

- QA irradiation should be conducted at less than -15°C to avoid the sample annealing and simulate the ATLAS ITk at the the normal run time. To meet this requirement:
 - Temperature and Humidity: -20°C (Peltier + water cooling), $<5\%\text{RH}$
 - Irradiation fluence: using foil activation technique





IHEP: QA site – CCE

- β -ray penetrating mini sensor detected by scintillator
- Output charge measured by the ALiBaVa system.
- Collected charge is defined as the Most Probable Value of Landau function obtained from the fit of Landau convoluted with Gaussian to distribution of measured cluster charges.
- Temperature and Humidity: -20°C (controlled by the refrigerator), $<10\%\text{RH}$ (Dry air flowed into the refrigerator)



IHEP: QA site — IV, CV

- To estimate the IV and CV characteristics of the sensors, the MD8 was attached to the ITK Strip TEG board PCB and placed in the chamber .
- Temperature and Humidity: 20°C (unirradiated), $\pm 20^\circ\text{C}$ (irradiated), controlled by the chamber , <10%RH (Dry air flowed into the climate chamber)




The MD8 was mounted on the PCB for IV and CV measurements


IV/CV/CCE Site Qualification

- Samples irradiated at the CSNS, with subsequent IV, CV, and CCE measurements completed at IHEP.
- Do the gamma spectroscopy and ship them to Birmingham and Ljubljana site for independent cross-verification.

To Birmingham




To Ljubljana



- The total α dose for all samples is $0 < 11.5(\text{cps})$, and the β/γ dose is $104.1 < 115(\text{cps})$, both meeting the standard.
- The sample is in line with expectations and has basically only Na^{22} radioactive source, with peaks of 0.511MeV and 1.27MeV observed as expected.


CCE test site qualification



- CCE results measured at IHEP showed good consistency with results from Ljubljana and Birmingham(difference: <10%) .


IV test site qualification

- The same dose of 1.6×10^{15} , the IVs at both sites are about 22 μA .
- The damage constant α is evaluated as $(3.83 \pm 0.22) \times 10^{-17}$ A/cm, consistent with the value of $\alpha_{80/60} = (4.06 \pm 0.04) \times 10^{-17}$ A/cm from the literature(ATL-ITK-PROC-2023-001)



CV test site qualification




- The full depletion voltage was around 220 V for un-irradiated MD8, meeting the QA requirement (no more than 350 V).
- Samples with similar fluence have close full depletion voltages.
- The curves of W5966 and W5958 are inconsistent.



Further investigation on CV

- In order to understand the reasons for these differences, we customized a new box according to the community's opinion to achieve the same test setup as other sites.


● Circuit connection:



- Disconnect the **LP LC-bias ring** connection during open calibration
- Connect the **LP LC-bias ring** during CV measurements



New box layout and connection



Further investigation on CV


- We did some tests to check whether it was caused by the different designs of decoupling.

Standard Capacitance

Capacitor	1	2	3
Old box(pF)	97.47	96.66	97.63
New box(pF)	98.19	97.5	97.93

- We have verified through experiments with standard capacitors that both the old and new boxes can measure correct results.



CV comparison on irradiated samples (10kHz)



- The results of new and old boxes are basically the same, with only 10% different in $1/C^2$

Further investigation on CV

Unirradiated Capacitance result (1kHz)




- At 20°C, the capacitance at full depletion for the old box (23.8 pF) is consistently about 4 pF higher than that of the new box (19.6 pF).
- At -10°C, the capacitance at full depletion for the old box (22.5 pF) is consistently about 4 pF higher than that of the new box (20.2 pF).

Further investigation on CV

Test comparison using SiC samples

- We borrowed a SiC sample from another group for a comparison test between the new and old boxes. This sample is designed with a full depletion depth of 50 μm , corresponding to a full depletion voltage of around 100V and a capacitance of approximately 7–8 pF. This sample has undergone cross-checking. (their official setup: 10kHz, 1V, CpRp)

SiC Capacitance




- The old box measurement is consistent with results from other frequency stations, showing a capacitance of approximately 7.7 pF after full depletion.
- In contrast, the new box measures around 4.5 pF at full depletion, resulting in an absolute discrepancy of about 3 pF compared to the old box.

Further investigation on CV

- We used test setup to scan range from 100Hz to 2MHz
- $-20^{\circ}\text{C}/-10^{\circ}\text{C}$, 250V/500V humidity $< 10\%$
- Not observe any clear region where the inductive behavior (i.e., impedance increasing with frequency) dominates, even up to the maximum frequency supported by the KEYSIGHT E4980A (2 MHz).
- There is a slight difference due to temperature in the capacitance-dominated region, and all four datasets show very good overlap in the high-frequency region (100k – 1M Hz).

$$Z = R + j\omega L + \frac{1}{j\omega C}$$

$$|Z| = \sqrt{R^2 + \left(\omega L - \frac{1}{\omega C}\right)^2}$$






Summary

- IHEP has passed the CCE/IV/CV site qualification and CSNS pass the irradiation site qualification.
- China ATLAS ITk group has fully prepared all materials and devices required for irradiation, and now only awaits experimental approval from the irradiation sites to commence operations.
- With the assistance of the community, we conducted extensive CV-related inspections, which have significantly enhanced our understanding of the post-irradiation CV results.

Thank you!




Low temperature irradiation setup

- QA irradiation should be conducted at less than -15°C to avoid the sample annealing and simulate the ATLAS ITk at the the normal run time. To meet this requirement:
- Temperature ($T \sim -15^{\circ}\text{C}$) control: Semiconductor refrigeration sheet + aluminum sheet
- Humidity ($\text{RH} \sim 5\%$) control: Air compressor + dryer
- Temperature monitoring: Thermocouple + Thermometer
- Humidity monitoring: Electronic hygrometer





IV and cv test on md8 diodes

- For unirradiated diodes, the current should not exceed $0.1 \mu\text{A}/\text{cm}^2$ at 500V (RH
- The diodes should show a leakage current of less than $0.1 \text{ mA}/\text{cm}^2$ at 500V (-20°C) after irradiation to $1.6 \times 10^{15} \text{ neq}/\text{cm}^2$.
- Onset of micro-discharge should be at $\text{VMD} > 500\text{V}$ for both unirradiated and irradiated diodes.
- For unirradiated diodes, the depletion voltage (VFD) should be less than 350V.





CCE test on mini sensors

- ALiBaVa Setup
- Landau convolutional Gaussian distribution of signal spectrum
- Obtain the ADC value of the Landau peak
- Combined with the calibration to get the collected charge
- At 500 V, the CCE of data points for unirradiated mini sensor $\sim 23\text{ke}^-$, for $1.0 \times 10^{15} \text{ neq/cm}^2 \sim 15\text{ke}^-$, for $1.6 \times 10^{15} \text{ neq/cm}^2 \sim 11\text{ke}^-$. The collaboration has established a minimum threshold of 7.5 ke^- .



IV/CV test site qualification



1kHz

- Samples are shipped to Birmingham and Ljubljana site for independent cross-verification
- The same dose of 1.6×10^{15} , the IVs at both sites are about 22 μA .
- The damage constant α is evaluated as $(3.83 \pm 0.22) \times 10^{-17}$ A/cm, consistent with the value of $\alpha_{80/60} = (4.06 \pm 0.04) \times 10^{-17}$ A/cm from the literature(ATL-ITK-PROC-2023-001)
- The full depletion voltage was around 220 V for un-irradiated MD8, meeting the QA requirement (no more than 350 V).